

JAN 13 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130294</u> ✓
A	Name <u>Ben Dudley</u>	ID# <u>3532</u> ✓ Date <u>1.7.11</u> ✓
	Agency <u>Wasilla PD</u>	Phone # <u>907.352.5408</u>
	Instrument Location <u>1800 E. Parks Hwy, Wasilla, AK (WPD)</u>	
B	Alco S/N <u>X172967</u> ✓	Target Value <u>.080</u> ✓ High Pressure <u>1000</u>
	Alco Test Value Average <u>.080</u> ✓	<u>.081</u> ✓
	1 st Alco <u>.081</u> ✓	2 nd Alco <u>.081</u> ✓
Signature <u>Ben Dudley</u>		(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of FEB, 2011 DN

Daniel L. Nicholas
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

130294

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

DIAGNOSTIC CHECK

2004 07 29:

2/8/14

-- BREATH ANALYSIS --

ALCO ADJUSTED FOR 30.37 in	0.001	15:33
ALCO TARGET	0.000	15:33
BLANK TEST	0.000	15:34
INTERNAL STANDARD	VERIFIED	15:34
ALCO TO 30.37 in	0.000	15:35
BLANK TEST	0.000	15:35
SUBJECT SAMPLE <i>5mb</i>	0.000	15:36
BLANK TEST	0.000	15:37
ALCO TO 30.37 in	0.001	15:37
BLANK TEST	0.000	15:37

JANUARY 27, 2001
THE 13TH

— DIAGNOSTIC CHECK —

```

COMPUTER:          OKAY
PROGRAM:            OKAY
SOFTWARE DATE:      01/08/89
HEATERS
SAMPLE CHAMBER:    45c
BREATH TUBE:        42c
BAROMETER:          30.37 in
FLOW DETECTOR:      OKAY
PUMP
HIGH SPEED:         OKAY
DETECTOR:           OKAY
FILTERS:             OKAY
QUARTZ STANDARD:    OKAY
CALIBRATION:         OKAY

```

[illegible]

```
-L'0123456789+-.,!@#$%^&*~`|_{}[]\;:'"/<>?
-abcdefghijklmnopqrstuvwxyzABCDEFGHIJKLMNOPQRSTUVWXYZ
-abcdefghijklmnopqrstuvwxyzABCDEFGHIJKLMNOPQRSTUVWXYZ
-abcdefghijklmnopqrstuvwxyzABCDEFGHIJKLMNOPQRSTUVWXYZ
```